

AUDIOMETER LEVEL TEST

Plant Louisville Date 3-23-72
 Audiometer Make Maico Model ----- Serial 7910
 ASA ISO X Ambient Level 58

This audiometer has been calibrated with a General Radio 1565-Z Audiometer Calibration Set, at a sound level reading corresponding to a 60 db hearing level.

Frequency - Audiometer	125	250	500	1000	2000	3000	4000	6000	8000
Initial Setting - Audiometer									
Reading - Red Earphone		84.8	71.0	69.2	70.8	67.5	67.5	65.0	72.2
Reading - Blue Earphone		83.5	70.1	68.4	69.8	65.0	66.2	64.9	70.7
Actual SLM Reading		85.0	70.7	67.8	70.0	66.6	66.5	64.9	73.9
Correction	Red Earphone								
	Blue Earphone								
		-0.2	+0.3	+1.4	+0.8	+0.9	+1.0	+0.1	-1.7
		-1.5	-0.6	+0.6	-0.2	-1.6	-0.3	0	-3.2

Remarks:

I certify that I have performed this calibration on this date.

Allowable variation at each frequency: ± 4 db for frequencies 125-2000; ± 5 db for frequencies of 3000 and above.