

Name: Hegwood
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-15-1994
Test time: 2:00:28 pm
Test subject last name: Hegwood
Test subject first name: Joseph
ID number: 426198535
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	17600	0.150	117000	PASS
DB	18050	0.470	38500	PASS
SS	19950	0.350	57000	PASS
UD	21600	0.490	44100	PASS
R	23850	3.230	7390	PASS
NB	25750	0.170	151000	PASS

OVERALL FIT FACTOR = 27690 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-15-94
Test Subject Joseph R. Hegwood Date 8/15/94