

Name: WORLOW
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 10-12-1994
Test time: 9:54:15 am
Test subject last name: WORLOW
Test subject first name: CHARLES
ID number: 426941635
Operator name: STEWART
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID
Size: MEDIUM
Model:
Manufacturer: 3M
Approval number: TC-21C-488

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	3575	0.310	11500	PASS
DB	3885	0.490	7930	PASS
SS	4140	0.210	19700	PASS
UD	4290	0.150	28600	PASS
R	4395	3.070	1430	PASS
NB	4475	1.700	2630	PASS

OVERALL FIT FACTOR = 4353 PASS
(Calculated per ANSI Z88.2-1980)

Operator Arthur J. Stewart Date 10-12-94
Test Subject Charles Worlow Date 10-12-94