

Name: Dobbs
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-15-1994
Test time: 9:31:40 am
Test subject last name: Dobbs
Test subject first name: Willis
ID number: 426767397
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Large
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	15950	0.170	94300	PASS
DB	17650	0.240	73700	PASS
SS	18650	0.360	51900	PASS
UD	19450	0.220	88600	PASS
R	20750	0.450	46100	PASS
NB	21400	0.110	194000	PASS

OVERALL FIT FACTOR = 73553 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-15-94

Test Subject Willis Dobbs Date 8-15-94