

Name: Jackson
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 12:51:07 pm
Test subject last name: Jackson
Test subject first name: Willie
ID number: 428924018
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Large
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	26200	0.140	187000	PASS
DB	25750	0.410	62900	PASS
SS	25650	0.430	59700	PASS
UD	25200	0.220	114000	PASS
R	25000	1.830	13700	PASS
NB	25150	0.080	314000	PASS

OVERALL FIT FACTOR = 48802 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94
Test Subject Willie Jackson Date 8-12-94