



Interoffice Communication

To R. B. Martin
From D. S. Cox
Date August 27, 1979
Subject CHECKING THE CALIBRATION OF THE "REAL" MINIMONITOR AND
REPRODUCIBILITY MEASUREMENT OF THE CARBON ANALYSIS

A program was developed to determine if the calibration of the "REAL" badge membrane would change; however, the work changed to more of a precision study for the carbon adsorption test.

SUMMARY

1. A series of badges were exposed to known concentrations of VCM in nitrogen for known times and the calibration values measured.
2. Carbon was exposed to VCM (without using the badge) and the precision measured. Ultrasonic treatment improved the precision by almost a factor of 10.
3. Carbon was exposed to VCM and analysis done by the OSHA approved carbon disulfide method. The precision was only marginally better.

CONCLUSIONS

1. The badge calibrations had changed indicating the membrane was partially blocked. Replicate measurements are needed to overcome the inherent variability of the carbon analysis.
2. The head space method gives equivalent results to the carbon disulfide method.
3. Under normal analysis methods, the precision (1 standard deviation) of the carbon analysis is 19% relative (1 ppm TWA-8 hr.). There is some indication that improvement is possible using an ultra sonic treatment.

EXPERIMENT I

The badges were exposed to known concentrations of flowing vinyl chloride gas. The exposures were made in a special chamber so each badge had its own atmosphere unaffected by any of the other units. The analysis was done by F-42 head space analysis using a mixture of dimethylacetamide (DMA) and water (see Appendix for analysis details). The "Standard" used was the average value of 10 knowns. Table I shows the results of the calibration checks.

Due to the poor relative standard deviation, a search was made for the source of the variability.

CC CLM
AHS
RAF
JVL

SAL 000048121

EXPERIMENT II

A systematic search for variability was made in the test. Six vials were filled with a 12 ppm VCM standard gas and run on the F-42 (Table 2). The errors due to the injection step were $\pm 2.7\%$ relative.

A solution of VCM in DMA was prepared and an equal amount of this solution put into five F-42 vials. The extraction solvent was added and the vials tested on the F-42 (Table 3). The precision was still $\pm 2.7\%$ relative and showed that the equilibrium of VCM in the level-space was constant.

EXPERIMENT III

The regular 90°C method of analysis of the carbon standards (Table 4) showed a relative error of 19.3%. A test was made at a lower bath temperature (70°C) to check for VCM degradation. None was evident. The relative error for standards in the 70°C bath was not significantly better.

EXPERIMENT IV

Carbon was put into a flask and exposed to VCM vapors then the carbon was put into vials (equal weights) and run. The precision was much better (Table 5) indicating standard preparation contributes to the errors.

EXPERIMENT V

A test was carried out with propane to confirm that the errors were not due to the reactivity of VCM. Table 6 shows the results of this test. Propane behaved the same as VCM on carbon.

EXPERIMENT VI

A set of standards were analyzed by the carbon disulfide method to confirm that the head space method was not the source of errors. Table 7 shows the results of that test. There was some improvement over the regular carbon analysis.

EXPERIMENT VII

A test was made using the ultrasonic bath to dislodge any gas bubbles from the carbon after extraction was complete.

The results (Table 8) were favorable with a relative error of only 3.5%. More tests will be needed to confirm that the ultrasonic treatment results were typical and not a fluke.

David S. Cox
David S. Cox
Mfg. Chemist

SAL 000048122

TABLE I
CALIBRATION VALUES

BADGE #	209	210	211	213	214	116
Run #1	1.01	1.05	1.17	1.29	1.09	1.71
2	1.00	1.01	1.15	0.93	1.03	1.00
3	1.03	1.04	0.90	1.00	1.02	1.02
4	0.89	0.90	0.94	0.90	0.98	1.04
5	0.93	0.92	0.88	0.80	0.89	0.75
6	1.47	1.48	0.87	1.47	1.66	-
7	1.25	1.30	1.23	1.58	1.60	-
8	1.41	1.28	1.09	-	1.18	1.48
9	1.96	1.79	1.86	-	2.44	2.52
10	1.43	1.10	-	-	1.36	1.38
11	1.17	0.81	1.34	-	1.21	1.42
12	<u>1.23</u>	<u>1.42</u>	<u>1.78</u>	<u>-</u>	<u>1.30</u>	<u>0.91</u>
Average	1.23	1.18	1.20	1.14	1.31	1.32
Rel. Std. Deviation	+ 10%	24%	29%	27%	32%	39%

$$K = \frac{CT}{m}$$

K = Calibration #

T = Hours

M = Microliters of VC on Carbon

C = Concentration of the VC in microliters per liter (ppm).

consequence days

no data due to carbon water

TABLE 2

ANALYSIS OF 12 PPM VC STANDARD GAS

	<u>Area</u>
	2724
	2771
	2830
	2774
	2875
	<u>2662</u>
% Rel. Std. Deviation	2.7%

TABLE 3

REPLICATE ANALYSIS OF VCM SOLUTION

	2276
	2368
	2296
	2208
	<u>2230</u>
% Rel. Std. Deviation	2.7%

SAL 000048124

TABLE 4

ANALYSIS OF VC INJECTED INTO A VIAL CONTAINING CARBON

(10 μ l VC)

<u>Water Bath 90° for 2 Hrs.</u>			<u>Bath 70° for 17 Hrs.</u>
<u>#1</u>	<u>#2</u>	<u>#3</u>	
10.7	8.1	12.5	10.4
13.2	14.4	9.6	13.2
7.4	10.8	11.9	10.6
8.3	8.2	10.6	10.0
11.9	9.3	11.9	10.6
9.4	9.9	7.0	8.8
10.1	7.9	8.8	10.6
8.9	10.7	12.1	8.0
	9.9	9.0	6.9
	10.8	6.5	% Rel. Std.
COMBINED % REL. STD. DEV. 19.3%			Deviation 17.5%

TABLE 5

ANALYSIS OF CARBON EXPOSED TO VC IN A GLASS FLASK

(Equal Weights of Carbon)

<u>Area</u> <u>@ 90° for 1 Hr.</u>		<u>Area</u> <u>@ 70° for 2 Hr.</u>
1006		909
995		1021
1028		925
1003		917
973		1095
962		947
1031		797
866		1006
894		1032
966		1023
% Rel. Std.	5.6%	8.8%
Deviation		

SAL 000048125

TABLE 6

ANALYSIS OF PROPANE ON CARBON

	2.18
	2.20
	3.30
	3.89
	4.14
	2.58
	4.34
	<u>3.32</u>
% Rel. Std. Deviation	26.3%

TABLE 7

ANALYSIS OF CARBON WITH 10 ~~μ~~l VC BY

CARBON DISULFIDE METHOD

9.1	17.9
9.6	10.5
8.8	7.9
13.8	9.2
9.2	16.1
7.7	8.8
8.0	8.8
8.4	8.2
8.8	9.0

10% Rel. Std. Deviation

SAL 000048126

TABLE 8

ANALYSIS OF CARBON BY F-42 AFTER ULTRASONIC TREATMENT

10 nl STD ON CARBON (DMA EXTRATANT)

	<u>AREA</u>
	3380
	3598
	3477
	3509
	3368
	3494
	3745
	<u>3426</u>
% Rel. Std. Deviation	3.5%

SAL 000048127