cc: C. L. Bradford
P. L. Slayton - 1650

October 6, 1971

Mr. David O. Thompson
Sprague Electric Company
Engineering Laboratories
North Adams, Massachusetts 01248

Dear Mr. Thompson:

This letter will confirm our telephone conversation of this morning concerning Santovac 2 vacuum pump fluid. Please pass this information along to those who are using Santovac 2. Let me know if I can be of further help.

As indicated on their labels, Santovac 1 and Santovac 2 contain polychlorinated biphenyls (PCBs) of various types. As you are aware, PCBs have increasingly come under the scrutiny of various governmental agencies and scientific laboratories. In some studies, certain PCBs have been found to be persistent, an environmental contaminant and possibly injurious to certain forms of bird, aquatic and animal life. Accordingly, it is important that PCBs be kept out of the food chain.

We want to remind your firm of the extreme care that must be taken by users of Santovac 1 and Santovac 2 fluids to prevent any entry into the environment through spills, leakage, use, disposal, vaporization or otherwise. Some specific applications where these fluids should definitely not be used are in vacuum systems involving food or food-related products, animal feeds, pharmaceutical products, packaging materials involving any of such items and buildings where any of these items are manufactured or stored.

If your vacuum system involves any of the mentioned applications, or if your system is such that entry into the environment cannot be prevented, you should replace Santovac 1 and 2 with non-PCB fluids immediately. Please contact your pump supplier for his recommended replacement product.

DSW 018153

STLCOPCB4004698

Used Santovac 1 and 2 should be returned to Monsanto Company, W. G. Krummrich Plant, Sauget, Illinois 62201, Attention: Supervisor Dept. 246 for proper disposal. This service is performed at cost, which is $0.03 per pound.

Very truly yours,

R. E. Hatton
Technical Manager
Hydraulics and Lubricants

/pep

DSW 018154

STLCOPCB4004699