11-4-60

0.01 ml (50% in benzene)

PROGRAM TEMPERATURE INSTRUMENT # 2

Material: Pantiwax or Poly Phenols.

310 °C. Temp. Detector Oven at 75 Volts Variac. Recorder # 3617
310 °C. Temp. Inj. Block at 127 Volts Variac.
295 °C. Temp. Column Block at 86 Volts Variac. H.W.D. # 2
304 °C. Temp. Tube Lines at 28 Volts Variac. 225 M.A.
°C. Temp. Exit Lines at Volts Variac.
4.[illegible] CM at 30 Psig Reference Side:
5.9 CM at 30 Psig Sample Side: Helium
Column Data:
2 - 1 meter 25% Apiezon oil L on Chromsorb W #117 + 118

Temperature Program: 150 °C. → 380 °C.
Column Voltage 40 → 113 Volts
1 RPM Drive
15 Minutes Hold:
3 Minutes Cool:

0067841

TOWOLDMON0056692