

DATE: WEEK OF 3/10/78
THROUGH 3/16/78

AF
B4
RCA

AVERAGE AUCM LEVELS - BLDGS. 1, 111, 115, AND 121

	G.C. READINGS							
	TOTAL	BETWEEN 1 AND		BETWEEN 3 AND		GREATER		
	# SPLS.	30 PPM INCLUSIVE		100 PPM INCLUSIVE		THAN 100 PPM		
		#	%	#	%	#	%	
BUILDING 1								
ALARM READINGS								
FLOOR A	1956	421	21.52	96	4.91	0	0.00	
FLOOR B	1956	802	41.00	184	6.34	1	0.05	
FLOOR C	1956	890	45.50	156	7.98	3	0.15	
TOTAL	5868	2113	36.01	376	6.41	4	0.07	
3 RD FLOOR CONTROL RM.	978	7	0.72	0	0.00	0	0.00	
VENT AIR ENTRY	978	9	0.92	3	0.31	0	0.00	
STAIRWELL	978	2	0.20	1	0.10	0	0.00	
BREATHING AIR	978	2	0.20	1	0.10	0	0.00	
% UPTIME FOR WEEK - 97.5%								
BUILDING 111								
ALARM READINGS								
FLOOR A	1000	618	61.80	235	23.5	1	0.10	
FLOOR B	2000	889	44.45	139	6.95	1	0.05	
FLOOR C	2000	710	35.50	104	5.20	3	0.15	
TOTAL	5000	2217	44.34	478	9.56	5	0.10	
3 RD FLOOR CONTROL RM.	1000	15	1.50	4	0.40	1	0.10	
VENT AIR ENTRY	1000	12	1.20	9	0.90	1	0.10	
BREATHING AIR	1000	3	0.30	1	0.10	0	0.00	
ROOF	1000	389	38.90	113	11.30	1	0.10	
% UPTIME FOR WEEK - 99.8%								
BUILDING 115								
	1000	198	19.80	34	3.40	1	0.10	
% UPTIME FOR WEEK - 99.8%								
BUILDING 121								
ALARM READINGS								
FLOOR A	1980	774	39.09	308	15.56	10	0.51	
FLOOR B	1980	1282	64.75	345	17.42	16	0.81	
FLOOR C	1980	1420	71.72	307	15.51	4	0.20	
TOTAL	5940	3476	58.52	960	16.16	30	0.51	
3 RD FLOOR CONTROL RM.	990	172	17.37	38	3.84	3	0.30	
VENT AIR INTAKE	990	26	2.63	20	2.02	2	0.20	
BEFORE BREATHING	990	28	2.83	11	1.11	0	0.00	
AIR FILTER								
AFTER BREATHING	990	103	10.40	12	1.21	0	0.00	
AIR FILTER								
% UPTIME FOR WEEK - 97.6%								

BY: REMOSSER

CC: GFL, SSM, CTU, HHS, HEB, RSK, STB, RET, LDP,
DS, CHH, TCP, MJC, WOF, HLH, ASK, JDE

NOTE: LPA AVG. AUCM LEVELS WILL BE DISTRIBUTED UPON RECEIPT
OF DATA.