

Name: Oliver
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-05-1994
Test time: 2:38:13 pm
Test subject last name: Oliver
Test subject first name: Gregory
ID number: 428231863
Operator name: J. Uptain
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113


Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	40300	0.100	403000	PASS
DB	40750	0.110	370000	PASS
SS	40000	0.200	200000	PASS
UD	39000	0.130	300000	PASS
R	38950	2.510	15500	PASS
NB	38900	0.060	649000	PASS

OVERALL FIT FACTOR = 75401 PASS
(Calculated per ANSI Z88.2-1980)

Operator  Date 8-5-94

Test Subject  Date 8-5-94