

Name: Smith  
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE  
Version 2.01 S/N 927  
VISTA POLYMERS

## FIT TEST REPORT

-----  
Test date: 08-10-1994  
Test time: 8:22:05 am  
Test subject last name: Smith  
Test subject first name: Erik  
ID number: 531768399  
Operator name: S. Carter  
Next fit test due:  
Test agent: Ambient particles  
Test device: TSI MODEL 8010 PORTACOUNT

## RESPIRATOR ID

Size: M  
Model: Ultravue  
Manufacturer: Facepiece  
Approval number: TC-19C-113

Additional notes:  
Additional notes:

## TEST DATA

-----  
Pass/Fail level: 100  
Test cycles.....: 1

| Ex. | Ambient<br>(Part/cc) | Mask<br>(Part/cc) | Fit Factor | Pass/Fail |
|-----|----------------------|-------------------|------------|-----------|
| NB  | 27950                | 0.250             | 111000     | PASS      |
| DB  | 30350                | 6.210             | 4890       | PASS      |
| SS  | 33450                | 19.300            | 1720       | PASS      |
| UD  | 38200                | 26.700            | 1420       | PASS      |
| R   | 40850                | 100               | 408        | PASS      |
| NB  | 41700                | 137               | 303        | PASS      |

OVERALL FIT FACTOR = 828 PASS  
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-10-94  
Test Subject Erik E. Smith Date 8/10/94

WANTS TO TRY SMALL SIZE