

Name: Allman
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-05-1994
Test time: 12:39:08 pm
Test subject last name: Allman
Test subject first name: Paul
ID number: 416705023
Operator name: J. Walker
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID
Size: Large
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	43350	1.210	35800	PASS
DB	44500	2.680	16600	PASS
SS	42700	1.090	39200	PASS
UD	41200	1.230	33500	PASS
R	42200	5.730	7370	PASS
NB	43950	1.410	31200	PASS

OVERALL FIT FACTOR = 19276 PASS
(Calculated per ANSI Z88.2-1980)

Operator

James Walker

Date

8-5-94

Test Subject

Paul Allman

Date

08-05-94