

Name: Bishop
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 1:50:37 pm
Test subject last name: Bishop
Test subject first name: Mickey
ID number: 427117304
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID
Size: Large
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	29450	0.100	294000	PASS
DB	29100	1.500	19400	PASS
SS	29000	0.090	322000	PASS
UD	28750	0.110	261000	PASS
R	28400	2.400	11800	PASS
NB	27700	0.120	231000	PASS

OVERALL FIT FACTOR = 39746 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94

Test Subject Mickey Bishop Date 8-12-94