

Name: CLAY
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 10-03-1994
Test time: 11:24:20 pm
Test subject last name: CLAY
Test subject first name: CHARLES
ID number: 427332440
Operator name: STEWART
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: LARGE
Model:
Manufacturer: 3M
Approval number: TC-21C-488

Additional notes:

Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	5350	0.370	14400	PASS
DB	6040	1.190	5070	PASS
SS	6070	0.740	8200	PASS
UD	5915	3.510	1680	PASS
R	5385	5.200	1030	PASS
NB	4565	1.750	2610	PASS

OVERALL FIT FACTOR = 2566 PASS
(Calculated per ANSI Z88.2-1980)

Operator

Arthur J. Stewart

Date

10-3-94

Test Subject

Clark R. Clay

Date

10-3-94